



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/539,572	06/17/2005	Bertrand Viellerohe	0501-1136	4337
22511	7590	05/17/2010	EXAMINER	
OSHA LIANG L.L.P. TWO HOUSTON CENTER 909 FANNIN, SUITE 3500 HOUSTON, TX 77010			WERNER, DAVID N	
			ART UNIT	PAPER NUMBER
			2621	
			NOTIFICATION DATE	DELIVERY MODE
			05/17/2010	ELECTRONIC

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

docketing@oshaliang.com

buta@oshaliang.com

Advisory Action
(Continuation of PTOL-303)

Applicant presents two arguments: first, that the microtip of *Gorecki* is not the "lens" as claimed, and second, that it is improper to combine the SNOM microscope of *Gorecki* with the confocal microscope of *Naruse* to reproduce the claimed confocal microscope.

First, regarding the "lens", it is respectfully submitted that since light is transmitted through the non-planar microtip, it is inherently a "lens" even if it does not have the classical lenticular shape as shown in figures 3, 5, and 6 of the present application.

Next, regarding the alleged improper combination, it is respectfully submitted that it was known in the art the time of the invention to use a VCSEL laser for both confocal and SNOM applications. See for example Göhde et al., "A Low Temperature Scanning Confocal and Near-Field Optical Microscope" (June 1997), describing a microscope which can be set up with interchangeable tips either for confocal or SNOM operation.

It is noted that Applicant states, but does not claim, that a difference between the present invention and *Gorecki* is that in SNOM systems like *Gorecki*, the microtip is placed at the nanometer scale from the objective, whereas in VCSEL systems like the present invention, the lens is placed at the micrometer to millimeter scale from the

Art Unit: 2621

objective. If placed in the claims, this limitation would overcome the *Goreck* reference as prior art.

/Mehrdad Dastouri/

Supervisory Patent Examiner, Art Unit 2621